

June 16, 2005

Mr. Harry H. Hodes
Acme Testing Co.
P.O. Box 3,
2002 Valley Highway
Acme, WA 98220-003
USA

Dear Mr. Hodes,

The Bureau has received a renewal submission for your 3m & 10m Open Area Test Site #2 located in Acme, WA, USA. I have reviewed the report and find it complies with RSS 212, Issue 1 (Provisional).


The site remains acceptable to Industry Canada for the performance of radiated measurements. Please continue to reference the file number assigned to the site (**IC 3251**) in the body of all test reports containing measurements made on this site.

This reference number is the indication to the Industry Canada Certification Officers that the site meets the requirements of RSS 212, Issue 1 (Provisional). Your company has been added to our published list of filed sites on the Bureau's web page. Please keep the contact information current by notifying us if it changes or is in error.

Keep informed of the latest Industry Canada regulations by visiting the Bureau's site on the World Wide Web: <http://strategis.ic.gc.ca/certification>

Whenever major construction or repairs to the site are completed, a re-submission of the site attenuation characteristics will be required, or every three years.

Yours sincerely,



Stéphane Proulx
For Wireless Laboratory Manager
Certification and Engineering Bureau
3701 Carling Ave., Building 94
P.O. Box 11490, Station "H"
Ottawa, Ontario
K2H 8S2
Tel. No. (613) 990-3796
Fax. No. (613) 990-4752